



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	1 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
Permanent Facility					
1	CLINICAL BIOCHEMISTRY	Blood	Hb Electrophoresis- Detection of Hb A, Hb S and Hb F	agarose gel PH 8.6,TEB buffer	NA
2	CLINICAL BIOCHEMISTRY	Fluid	ADA	Berthelot	20.0
3	CLINICAL BIOCHEMISTRY	Fluid	Glucose	GOD-POD End point	4.5
4	CLINICAL BIOCHEMISTRY	Fluid	Protein	Biuret	3.4
5	CLINICAL BIOCHEMISTRY	Serum	ADA	Berthelot	12.1
6	CLINICAL BIOCHEMISTRY	Serum	Bilirubin Indirect	Calculated	NA
7	CLINICAL BIOCHEMISTRY	Serum	Corrected Calcium	Calculated	NA
8	CLINICAL BIOCHEMISTRY	Serum	eGFR	Calculated - Modified MDRD equation	NA
9	CLINICAL BIOCHEMISTRY	Serum	Ketone bodies	Qualitative nitroprusside	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 2 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
10	CLINICAL BIOCHEMISTRY	Serum	LDL Cholesterol	Calculated	NA
11	CLINICAL BIOCHEMISTRY	Serum	Osmolality	Calculated	NA
12	CLINICAL BIOCHEMISTRY	Serum	Protein Electrophoresis- Detection of M band	Agarose pH 8.6 Veronal Buffer	NA
13	CLINICAL BIOCHEMISTRY	Serum	TIBC	Calculated	NA
14	CLINICAL BIOCHEMISTRY	Serum	VLDL Cholesterol	Calculated	NA
15	CLINICAL BIOCHEMISTRY	Urine	Calcium	Calculated as mg/day from 24 hours urine	NA
16	CLINICAL BIOCHEMISTRY	Urine	Calcium/Creatinine Ratio	Calculated	NA
17	CLINICAL BIOCHEMISTRY	Urine	Creatinine Clearance	Calculated	NA
18	CLINICAL BIOCHEMISTRY	Urine	Ketone bodies	Qualitative nitroprusside	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	3 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
19	CLINICAL BIOCHEMISTRY	Urine	Na ⁺	Direct ISE	1.6
20	CLINICAL BIOCHEMISTRY	Urine	Phosphorous	Ammonium Molybdate	4.7
21	CLINICAL BIOCHEMISTRY	Urine	Protein	Biuret	3.4
22	CLINICAL BIOCHEMISTRY	Urine	Protein	Calculated	NA
23	CLINICAL BIOCHEMISTRY	Urine	Uric Acid	Uricase POD	3.8
24	CLINICAL BIOCHEMISTRY	Blood	Dithionite Test for HbS	Dithionite	NA
25	CLINICAL BIOCHEMISTRY	Fluid	Protein	Pyrogallol red	14.4
26	CLINICAL BIOCHEMISTRY	Plasma	Glucose	GOD-POD End point	4.5
27	CLINICAL BIOCHEMISTRY	Serum	Albumin	BCG	4.3



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 4 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
28	CLINICAL BIOCHEMISTRY	Serum	Alkaline Phosphatase	pNPP with AMP buffer	5.1
29	CLINICAL BIOCHEMISTRY	Serum	ALT	L-Alanine LDH UV kinetic	6.3
30	CLINICAL BIOCHEMISTRY	Serum	Amylase	CNPG	5.0
31	CLINICAL BIOCHEMISTRY	Serum	Bilirubin Direct	Diazo Reaction	11.2
32	CLINICAL BIOCHEMISTRY	Serum	Bilirubin Total	Diazo Reaction	4.7
33	CLINICAL BIOCHEMISTRY	Serum	Calcium	Arsenazo III	5.0
34	CLINICAL BIOCHEMISTRY	Serum	Cholesterol Total	CHOD-POD	3.4
35	CLINICAL BIOCHEMISTRY	Serum	Cholinesterase	Butyrylcholinesterase	20.3
36	CLINICAL BIOCHEMISTRY	Serum	CK	Phosphocreatinine, HK, G6PD	7.1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	5 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
37	CLINICAL BIOCHEMISTRY	Serum	Creatinine	Jaffe two point	5.6
38	CLINICAL BIOCHEMISTRY	Serum	HDL Cholesterol	Dextran-Sulphate-Mg+2 Precipitation	8.9
39	CLINICAL BIOCHEMISTRY	Serum	Iron	Ferrozine	12.0
40	CLINICAL BIOCHEMISTRY	Serum	K+	Direct ISE	3.1
41	CLINICAL BIOCHEMISTRY	Serum	LDH	Lactate to Pyruvate, UV Kinetic	6.2
42	CLINICAL BIOCHEMISTRY	Serum	Na+	Direct ISE	1.6
43	CLINICAL BIOCHEMISTRY	Serum	Phosphorus	Ammonium Molybdate	4.7
44	CLINICAL BIOCHEMISTRY	Serum	Total Protein	Biuret	3.4
45	CLINICAL BIOCHEMISTRY	Serum	Triglycerides	Lipase GPO-POD	4.4



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 6 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
46	CLINICAL BIOCHEMISTRY	Serum	UIBC	Ferrozine - Alkaline pH	5.4
47	CLINICAL BIOCHEMISTRY	Serum	Urea	Urease GLDH	7.1
48	CLINICAL BIOCHEMISTRY	Serum	Uric Acid	Uricase POD	3.8
49	CLINICAL BIOCHEMISTRY	Urine	Calcium	Arsenazo III	5.0
50	CLINICAL BIOCHEMISTRY	Urine	Creatinine	Jaffe two point	5.6
51	CLINICAL BIOCHEMISTRY	Urine	Protein	Pyrogallol red	17.4
52	CLINICAL PATHOLOGY	Body Fluids	Microscopy Examination	Microscopy	NA
53	CLINICAL PATHOLOGY	Urine	Bile Salt	Manual- Hay's Sulphur Flower (Surface Tension) Test	NA
54	CLINICAL PATHOLOGY	Urine	Microscopic Examination	Microscopy	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	7 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
55	CLINICAL PATHOLOGY	Urine	Albumin	Protein Error Of Indicator- Multistrip Method	NA
56	CLINICAL PATHOLOGY	Urine	Bile Pigment	Fouchet Test - Manual / Multistrip Method	NA
57	CLINICAL PATHOLOGY	Urine	Glucose	Glucose Oxidase- Peroxidase Multistrip Method	NA
58	CLINICAL PATHOLOGY	Urine	Ketone Bodies	Rothera's Test (Nitroprusside Reaction) -Multistrip Method	NA
59	CLINICAL PATHOLOGY	Urine	Occult Blood	Peroxidase Like Activity of Hb- Multistrip Method	NA
60	CLINICAL PATHOLOGY	Urine	pH	Double Indicator Principle- Multistrip Method	NA
61	CLINICAL PATHOLOGY	Urine	Specific Gravity	Concentration of Ions- Multistrip Method	NA
62	CLINICAL PATHOLOGY	Urine	Urobilinogen	Ehrlich's Reaction- Multistrip Method	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 8 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
63	CYTOPAT HOLOGY	Body fluids (ascitic, pleural, cerebrospinal fluid, synovial, pus, bronchoalveolar lavage fluid, endotracheal tube secretion, nipple discharge)	Cytospin/ Centrifugation, Smear Preparation, PAP, H&E and MGG Staining. Light microscopy with interpretation	As per laid down SOP's.	NA
64	CYTOPAT HOLOGY	Cervical and vaginal smears (conventional)	PAP Stain. Light microscopy with interpretation	As per laid down SOP's.	NA
65	CYTOPAT HOLOGY	Palpable or nonpalpable lesions involving any organ	FNA, Smear Preparation, PAP, H&E and MGG Staining. Light microscopy with interpretation	As per laid down SOP's.	NA
66	CYTOPAT HOLOGY	Scrapings/ brushings (gastrointestinal tract, bronchial, oral)	PAP, H&E stain and MGG Stain. Light microscopy with interpretation	As per laid down SOP's.	NA
67	CYTOPAT HOLOGY	Unstained smear	PAP, H & E Stain and MGG Stain. Light microscopy with interpretation	As per laid down SOP's.	NA
68	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Peripheral Smear Examination	Microscopy (Giemsa Stain)	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 9 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
69	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Reticulocyte Count	Reticulocyte Stain (New Methylene Blue Stain/ Microscopy)	5.5
70	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Differential WBC Count(Basophil)	Electrical Impedance and Microscopy	3.5
71	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Differential WBC Count(Eosinophil)	Electrical Impedance and Microscopy	8.2
72	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Differential WBC Count(Lymphocyte)	Electrical Impedance and Microscopy	4.9
73	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Differential WBC Count(Monocyte)	Electrical Impedance and Microscopy	12.1
74	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Differential WBC Count(Neutrophils)	Electrical Impedance and Microscopy	2.8



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 10 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
75	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	ESR	Westergren method	13.4
76	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Hb estimation	Non- cyanmeth Hb estimation	2.4
77	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	HCT	Pulse Height Detection Method	2.9
78	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Identification and Quantification of HbA2	Ion Exchange High Performance Liquid Chromatography	8.6
79	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Identification and Quantification of HbF	Ion Exchange High Performance Liquid Chromatography	14.6
80	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Malarial Antigen Detection Test	Immunochemistry (Pf-HRP-2, P. vivax specific pLDH and Pan malaria specific pLDH)	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	11 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
81	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	MCH	Calculated	2.3
82	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	MCHC	Calculated	2.6
83	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	MCV	Calculated	1.7
84	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Platelet Count	Electrical Impedence	9.5
85	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	PSMP	Microscopy (Giemsa stain)	NA
86	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	RBC Count	Electrical Impedence	2.1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 12 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
87	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	RDW	Calculated	2.9
88	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Anti- coagulated blood(EDTA)	Total WBC Count	Electrical Impedance	4.3
89	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Blood (Citrated Plasma)	Activated Partial Thromboplastin Time	Manual Method/ Optical detection	5
90	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Blood (Citrated Plasma)	Prothrombin Time	Manual Method/ Optical detection	5
91	HAEMAT OLOGY & IMMUNOH AEMATOL OGY	Bone Marrow Material	Bone Marrow Smear Examination	Microscopy (Giemsa Stain)	NA
92	HISTOPA THOLOGY	Paraffin blocks for second opinion	Microtomy H & E Staining	As per laid down SOP'S Light Microscopy with Interpretation	NA
93	HISTOPA THOLOGY	Slides for second opinion	Stained/ Unstained slides	Light Microscopy with Interpretation	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA

Accreditation Standard ISO 15189:2012

Certificate Number MC-3017(in lieu of M-0450) Page No. : 13 / 17

Validity 31/10/2018 to 30/10/2020 Last Amended on 03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
94	HISTOPA THOLOGY	Small, medium & large tissues received in formalin	GrossingDecalcification , Processing, Paraffin EmbeddingMicrotomyH & E Staining	As per laid down SOP'S Light Microscopy with Interpretation	NA
95	HISTOPA THOLOGY	Tissue received in normal saline	Grossing- Processing & section cutting by cryostatH & E staining	As per laid down SOP'S Light Microscopy with interpretation	NA
96	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Alcian Blue Stain	As per laid down SOP'S Light Microscopy with Interpretation	NA
97	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Giemsa stain (May Gruwald Giemsa)	As per laid down SOP'S Light Microscopy with Interpretation	NA
98	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Grocott Methanamine Sliver Stain	As per laid down SOP'S Light Microscopy with Interpretation	NA
99	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Masson Trichome Stain	As per laid down SOP'S Light Microscopy with Interpretation	NA
100	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	PAS Stain (Periodic Acid Schiff)	As per laid down SOP'S Light Microscopy with Interpretation	NA
101	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Perl's Prussian Blue stain, for Iron	As per laid down SOP'S Light Microscopy with Interpretation	NA
102	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Reticulin Stain (Gordan and Sweet"s method)	As per laid down SOP'S Light Microscopy with Interpretation	NA
103	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	Toludine Blue Stain	As per laid down SOP'S Light Microscopy with Interpretation	NA
104	HISTOPA THOLOGY	Tissue/ paraffin block/ unstained slides	ZN (Zeihl Neesen stain for AFB)	As per laid down SOP'S Light Microscopy with Interpretation	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	14 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
105	MICROBIOLOGY & SEROLOGY	serum	Antibodies to HIV	ELISA	8.7
106	MICROBIOLOGY & SEROLOGY	serum	HBsAg	Rapid immunochromatography test	NA
107	MICROBIOLOGY & SEROLOGY	serum	HBsAg	ELISA	5.8
108	MICROBIOLOGY & SEROLOGY	serum	IgM Antibodies to Leptospira	ELISA	5.0
109	MICROBIOLOGY & SEROLOGY	serum	Widal Test	Tube agglutination test	NA
110	MICROBIOLOGY & SEROLOGY	Any clinical sample (Urine/pus/blood/body fluids/CSF/swab/sputum/tip/drain/stool/corneal rim/bile/BAL/scaring material/tissue/corneal button)	Acid fast stain	ZN stain method	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	15 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
111	MICROBIOLOGY & SEROLOGY	Any clinical sample (Urine/pus/blood/body fluids/CSF/swab/sputum/tip/drain/stool/corneal rim/bile/BAL/scaring material/tissue/corneal button)	Aerobic culture and sensitivity	Manual	NA
112	MICROBIOLOGY & SEROLOGY	Any clinical sample (Urine/pus/blood/body fluids/CSF/swab/sputum/tip/drain/stool/corneal rim/bile/BAL/scaring material/tissue/corneal button)	Gram stain	Microscopy	NA
113	MICROBIOLOGY & SEROLOGY	serum	Antibodies to HIV	Rapid dot immunoassay test	NA
114	MICROBIOLOGY & SEROLOGY	serum	Antibodies to HIV	Rapid immunoconcentration test	NA
115	MICROBIOLOGY & SEROLOGY	serum	Antibodies to HIV	Rapid immunochromatography test	NA
116	MICROBIOLOGY & SEROLOGY	serum	ASO	Latex agglutination test	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	16 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
117	MICROBIOLOGY & SEROLOGY	serum	CRP	Latex agglutination test	NA
118	MICROBIOLOGY & SEROLOGY	serum	Enterocheck-WB	Rapid immunochromatographic test	NA
119	MICROBIOLOGY & SEROLOGY	serum	HAV	Rapid immunochromatography test	NA
120	MICROBIOLOGY & SEROLOGY	serum	HAV	ELISA	5.8
121	MICROBIOLOGY & SEROLOGY	serum	HCV	ELISA	5.3
122	MICROBIOLOGY & SEROLOGY	serum	HCV	Rapid immunochromatography test	NA
123	MICROBIOLOGY & SEROLOGY	serum	HEV	Rapid immunochromatography test	NA



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name	NEW CIVIL HOSPITAL SURAT LABORATORY SERVICES, NEW CIVIL HOSPITAL, MAJURA GATE., SURAT, GUJARAT, INDIA		
Accreditation Standard	ISO 15189:2012		
Certificate Number	MC-3017(in lieu of M-0450)	Page No. :	17 / 17
Validity	31/10/2018 to 30/10/2020	Last Amended on	03/11/2018

S.No	Discipline	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used	%CV
124	MICROBIOLOGY & SEROLOGY	serum	HEV	ELISA	5.2
125	MICROBIOLOGY & SEROLOGY	serum	IgM Antibodies to Leptospira	Rapid immunochromatography test	NA
126	MICROBIOLOGY & SEROLOGY	serum	MAT for Leptospira	MAT	NA
127	MICROBIOLOGY & SEROLOGY	serum	RA	Latex agglutination test	NA
128	MICROBIOLOGY & SEROLOGY	serum	RPR	Slide Flocculation test	NA
129	MICROBIOLOGY & SEROLOGY	stool	Vibrio cholera	Hanging drop method	NA
130	MICROBIOLOGY & SEROLOGY	Throat swab	Albert stain	Albert stain	NA